



TECHNICAL STANDARDS &
SAFETY AUTHORITY
14th Floor, Centre Tower
3300 Bloor Street West
Toronto, Ontario
Canada M8X 2X4

Show facsimile of manufacturer's logo or trademark, as it will appear on the fitting, in the space below:

Endress+Hauser
EH

STATUTORY DECLARATION
Registration of Fittings

I, MIKE YOUROSKI, QUALITY MANAGER
(Name and Position, e.g. President, Plant Manager, Chief Engineer)

of ENDRESS+HAUSER CONDUCTA INC.
(Name of Manufacturer)

Located at 4123 LA PALMA AVE., ANAHEIM, CA, USA, 92807 1-714-577-5600
(Plant Address) (Telephone No.) (Fax No.)

do solemnly declare that the fittings listed hereunder, which are subject to the *Technical Standards and Safety Act*, Boilers and Pressure Vessels Regulation, comply with all of the requirements of ASME SECTION VIII, DIVISION 1, ASME B31.3
(Title of recognized North American Standard)

which specifies the dimensions, materials of construction, pressure/temperature ratings, identification marking the fittings and service;

or are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with _____ as supported by the attached data which identifies the dimensions, material of construction, pressure/temperature ratings and the basis for such ratings, the marking of the fitting for identification and service.

I further declare that the manufacture of these fittings is controlled by a quality system meeting the requirements of ISO 9001:2008 which has been verified by the following authority, SQS

The items covered by this declaration, for which I seek registration, are category H Fittings - CPA450-C*** type fittings. In support of this application, the following information and/or test data are attached as follows:
DRAWING EH-CRN-4, COMPONENT DRAWINGS, DESIGN CALCULATIONS
(drawings, calculations, test reports, etc.)

Declared before me at _____ in the _____ of _____
the _____ day of _____ AD 20_____.

Commissioner for Oaths:

(Printed name)
Michael J. ...
(Signature)

Mike Youroski
(Signature of Declarer)

FOR OFFICE USE ONLY

To the best of my knowledge and belief, the application meets the requirements of the *Technical Standards and Safety Act*, Boilers and Pressure Vessels Regulation, and CSA Standard B51 and is accepted for registration in Category H.

CRN: OH17834-5

NOTE: SEE ATTACHED
PART OF CRN
FOR THE SCOPE

REGISTERED

Registered by: Charley Dong

OH17834-5

Dated: July 23/15

Charley Dong

NOTE: This registration expires on July 22, 2025

July 23/15

ENDRESS+HAUSER CONDUCTA INC.
 4123 LA PALMA AVE.
 ANAHEIM, CA
 92807, USA

13-Jul-15

THIS IS PART OF
CRN 0117834-S
 Technical Standards & Safety Authority
 Bolts & Pressure Vessels
 Safety Division

SCOPE OF CRN REGISTRATION

PRODUCT DESCRIPTION

Description	Design Standard	Size Range	End Connection	Drawing	EH TI Document	NOTES
CPA450 Retractable Assembly	ASME B31.3, ASME VIII-1 UG-101(m)	1-1/4", 1-1/2", 2"	NPT, G Thread, ASME B16.5 Flanged	EH-CRN-4 and others	TI183C/07/EN/13.11	Note 1, Note 2

PRESSURE - TEMPERATURE RATINGS

MDMT	MAWP AT 100 F	MAWP AT MAWT	NOTES
Dependent on Material	340 psig at 100 F	340 psig at 300 F	Pressure-Temperature ratings limited by ASME B16.5 flange ratings when supplied with flanged connections.

MATERIALS OF CONSTRUCTION

STANDARD MATERIAL	NOTES
Type 316L Stainless Steel UNS S31603	Optional materials may be used. See Note 3.

MANUFACTURING LOCATIONS

Endress+Hauser Conducta Gesellschaft für Messund Regeltechnik mbH+Co.KG 70839 Gerlingen, Germany	Endress+Hauser Conducta Inc. 4123 LA Palma Ave., Anaheim, CA., USA, 92807
Endress+Hauser Analytical Instruments Suzhou Co. Ltd. China	

Note 1: When supplied, flanges to be in dimensional accordance with ASME B16.5. Non-integral flanges to be supplied with current CRN number.

Note 2: Various immersion depths are available. The immersion depth has no effect on the CRN Number.

Note 3: Optional materials of construction are: 316 S.Steel UNS S31600, Alloy 20 UNS N08020, Hastelloy C4 UNS N06455, Hastelloy C-276 UNS N10276, Hastelloy C22 UNS N06022, Hastelloy B3 UNS N10675, Super Duplex UNS S32750, Super Duplex UNS S32760, Titanium Grade 2 UNS R50400, Titanium Grade 3 UNS R50550.