



505 - 6th Street, Suite 200  
New Westminster, BC V3L 0E1

Toll Free: 1-866-566-SAFE  
Fax: (778) 396 - 2064  
www.safetyauthority.ca

PRESSURE VESSEL ENGINEERING  
B-120 RANDALL DRIVE  
WATERLOO ON N2V 1C6

**Date:** October 30, 2013  
**Account #:** 45363  
**Journal #:** 58285  
**Our File #:** 5473605

**Attn:** CATHEINE DIPLOCK

**Re:** Application for Design Registration

The design, as detailed in your, PVE-6396, for a Fitting is accepted for registration as follows:

**Registered To:** ENDRESS & HAUSER CONDUCTA CRN: 0F15740.51  
INC.

**MDMT:** 32 deg F

**MAWT:** 266 deg F

**MAWP:** 310 psig

**Drawing #:** 40236 & 402296

**Drawing Revision:** A, B

**Conditions Of Registration:**

Registration of OUA 260 Flowcell; 1" & 2"

This design was registered based on a technical review performed by the province of initial registration in accordance with the Association of Chief Inspectors policy on reciprocal recognition of design review.

**Reviewer's Notes:**

As required by CSA B51 4.2.1, this registration expires on January 8, 2023. This CRN is valid until the expiry date as long as the Manufacturer maintains a valid quality control program verified by an acceptable third-party agency until that date. Should the certification of the quality control program lapse before the expiry date, this registration shall become void.

Contact me if you have any questions. The invoice for registration will be forwarded under separate cover.

SHARON PETERS  
(778) 396-2027  
Sharon.Peters@safetyauthority.ca  
Design Administration

**cc:**