



TECHNICAL STANDARDS &
SAFETY AUTHORITY
14th Floor, Centre Tower
3300 Bloor Street West
Toronto, Ontario
Canada M8X 2X4

Show facsimile of manufacturer's logo or trademark, as it will appear on the fitting, in the space below

Endress+Hauser



STATUTORY DECLARATION

Registration of Fittings

I, Mike Yourroski, Quality Manager

(Name and Position, e.g. President, Plant Manager, Chief Engineer)

of Endress+Hauser Conducta Inc

(Name of Manufacturer)

Located at 4123 E. La Palma Ave., Suite 200, Anaheim, CA 92807-1813, USA +1714-577-5600 +1714-577-5672

(Plant Address)

(Telephone No.)

(Fax No.)

☐ do solemnly declare that the fittings listed hereunder, which are subject to the **Technical Standards and Safety Act**, Boilers and Pressure Vessels Regulation, comply with all of the requirements of

(Title of recognized North American Standard)

which specifies the dimensions, materials of construction, pressure/temperature ratings, identification marking the fittings and service;

☒ or are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with ASME B31.3 as supported by the attached data which identifies the dimensions, material of construction, pressure/temperature ratings and the basis for such ratings, the marking of the fitting for identification and service.

I further declare that the manufacture of these fittings is controlled by a quality system meeting the requirements of ISO 9001:2008

which has been verified by the following authority, SQS

The items covered by this declaration, for which I seek registration, are category F type fittings. In support of this application, the following information and/or test data are attached as follows:

PVEcl-6396, ISO Cert., Drw. no# 402296, 402312, 401654, 402306, 402317, 402312, Burts test reports, Design cover page

(drawings, calculations, test reports, etc.)

Declared before me at SEE ATTACHED JURAT in the _____ of _____
DATED 11/14/2012
the _____ day of _____ AD 20____.

Commissioner for Oaths:

ANHDAC VU
(Printed name)

(Signature)

Mike Yourroski

(Signature of Declarer)

FOR OFFICE USE ONLY

To the best of my knowledge and belief, the application meets the requirements of the **Technical Standards and Safety Act**, Boilers and Pressure Vessels Regulation, and CSA Standard B51 and is accepted for registration in Category F

CRN:

0F15740-5

NOTE: SEE ATTACHED

'PART OF CRN'

Registered by:

Charley Peng

FOR THE SCOPE

Dated:

Jan. 8/13

CSA
ISSA
Jan. 8/13

REGISTERED

CRN 0F15740-5

Charley Peng

Jan. 8/13

NOTE: This registration expires on Jan. 8, 2023

Technical Information for OUA260 Flowcell (NPT) Assemblies

The following information is for the range of flow sensors:

Design Code: ASME B31.3 2010 Ed.

Model and Sizes: OUA260, 1" and 2"

Drawing No: 402306 & 402296

Design pressure: 310psi

Design temperature: 266°F

MDMT: 32°F

Corrosion: None

Radiography: N/A

Hydrotest: 403psi @ 70°F (at assembly)

Body materials: A-182-316L/316 & B-164 (Monel 400)

Flange Head Screws: 18/8 (304) Class 2 Stainless Steel (UNS S30400)

