



505 - 6th Street, Suite 200  
New Westminster, BC V3L 0E1

Toll Free: 1-866-566-SAFE  
Fax: (778) 396 - 2064  
www.safetyauthority.ca

KEDKEP CONSULTING INC  
2296 LYNDHURST DRIVE  
OAKVILLE ON L6H 7V5

**Date:** March 3, 2017  
**Account #:** 51987  
**Journal #:** 67469  
**Our File #:** 5612367

**Attn:** CHARLES LIU

**Re:** Application for Design Registration

The design, as detailed in your, Model L & V Flow Meters, for a Fitting is accepted for registration as follows:

**Registered To:** KYTOLA INSTRUMENTS OY      **CRN:** 0F10611.21

**MDMT:** -20 deg F

**Drawing #:** 42414etc, Proof Tests, Tech Bull

**Drawing Revision:** a

**Conditions Of Registration:**

Registration of Cat. Model L & Model V Variable Area Flowmeters.

This design was registered based on a technical review performed by the province of initial registration in accordance with the Association of Chief Inspectors policy on reciprocal recognition of design review.

**Reviewer's Notes:**

Model L - MAWP: 239 psig, MAWT: 176 Deg F. Model V- MAWP: 266 psig, MAWT: 248 Deg F.

As required by CSA B51 4.2.1, this registration expires on MaY 20, 2020.. This CRN is valid until the expiry date as long as the Manufacturer maintains a valid quality control program verified by an acceptable third-party agency until that date. Should the certification of the quality control program lapse before the expiry date, this registration shall become void.

Contact me if you have any questions. The invoice for registration will be forwarded under separate cover.

SHARON PETERS

boiler.designregistration@safetyauthority.ca  
Design Administration

**cc:**