

---

TECHNICAL STANDARDS & SAFETY AUTHORITY  
345 CARLINGVIEW DRIVE  
TORONTO ON M9W 6N9

**Date:** March 7, 2019  
**Account #:** 35231  
**Journal #:** 72734

**Attn:** TANYA FRANCIS

**Re:** Application for Design Registration

The design, as detailed in your, TSSA SR# 2496258, for a Fitting is accepted for registration as follows:

**Registered To:** KYTOLA INSTUMENTS LTD      **CRN:** 0F21032.51

**Drawing #:** Scope of Registration

**Drawing Revision:** N/A

**Conditions Of Registration:**

Registration of Kytola Variable Are Flow Meters Model VLA, VEA & VDA as per attached SOR.

This design was registered based on a technical review performed by the province of initial registration in accordance with the Association of Chief Inspectors policy on reciprocal recognition of design review.

**Reviewer's Notes:**

As required by CSA B51 4.2.1, this registration expires on January 25, 2029. This CRN is valid until the expiry date as long as the Manufacturer maintains a valid quality control program verified by an acceptable third-party agency until that date. Should the certification of the quality control program lapse before the expiry date, this registration shall become void.

Any additional conditions of registration stated in TSSA CRN# 0F21032.5 registration shall apply to BC registration.

Contact me if you have any questions. The invoice for registration will be forwarded under separate cover.

Emilia Tam

emilia.tam@technicalsaftybc.ca  
Design Administration

**cc:**