

Direction générale de l'inspection

Montréal, 31 octobre 2016.

MR. TONY COLLINS PRIORCLAVE LTD 129-131 NATHAN WAY,WOOLWICH LONDON U.K. SE28 0AB

Manufacturer : PRIORCLAVE LTD 129-131 NATHAN WAY,WOOLWICH LONDON U.K. SE28 0AB

 OUR REFERENCE : 955755

 Design number :
 D-QCA-100V-ST-ASM-000010 REV. D-03 SHT 1 TO 3

 D-QCA-150V-ST-ASM-000010 REV. C-03 SHT 1TO 3

Subject: Registration of designs - Confirmation of approval

Dear Sir,

The Régie du bâtiment du Québec (RBQ) informs you that your application for the registration of designs has been evaluated and that the design was registered the 26-10-2016 under the Canadian Registration Number (CRN\NEC) : **E02229.6.**

We bring your attention to certain requirements of the regulation respecting pressure vessels, and the codes and standards adopted :

- The manufacturer must maintain a valid quality control program to produce according to the CRN.
- The manufacturer shall submit to the RBQ a copy of Manufacturer's Data Report for equipment manufactured according to this CRN within 30 days following the signing of this Declaration.
- The drawing number, including the revision number registered under this CRN, must be indicated on the Manufacturer's Data Report for equipment manufactured according to the CRN.

This notice of approval does not relieve the constructor of his responsibilities with respect to the design or construction of equipment manufactured according to this CRN.

Yours sincerely,

Bureau d'expertise et d'homologation en équipements sous pression

Montréal 545, boul. Crémazie Est, 7ième étage Montréal (Québec) H2M 2V2 Téléphone : 514 873-6459 Sans frais : 1 866 262-2084 www.rbq.gouv.qc.ca