

December 08, 2020

**Attention:** Scott Islip  
ROUND ENGINEERING INC  
10 SEGWUN ROAD  
WATERDOWN, ON L8B 0K6

The design submission, tracking number 2020-04824, originally received on October 02, 2020 was surveyed and accepted for registration as follows:

**CRN :** Z3457.2

**Accepted on:** December 08, 2020

**Reg Type:** NEW DESIGN

**Drawing No. :** D-QCA-150H-ST-ASM-000010 SHEETS 1-4 Rev Y-01

Design registered in the name of : PRIORCLAVE LTD

Description	MAWP	Design Temperature	MDMT
Internal Pressure	240kPa	138 °C	-10 °C
External Pressure	103kPa	138 °C	-10 °C

**The registration is conditional on your compliance with the following notes:**

*Acceptance is based on the understanding that the depth of the thread on the pressure door is not less than the diameter of the bolts for locking arm .*

*The tolerances between the pressure door and the locking arm are not part of this registration , also the components for the final locking arm lock are not part of this registration.*

*This vessel may not be pressurized when its temperature is colder than -10 [deg.C]. This note to appears on MDR*

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

If you have any question don't hesitate to contact me by phone at (780) 433-0281 ext 3335 or fax (780) 437-7787 or e-mail [Dedovic@absa.ca](mailto:Dedovic@absa.ca).

Sincerely,



DEDOVIC, BLAZO, M. Eng.  
DOP Cert. No. D00005642