

N-CUBED ENGINEERING
#409-8 MAIN STREET S
HAMILTON ON L8N 1E8**Date:** June 5, 2018
Account #: 58794
Journal #: 70521**Attn:** CUNG NGUYEN**Re: Application for Design Registration**

The design, as detailed in your, Renewal of CRN 0F12048.51, for a Fitting is accepted for registration as follows:

Registered To: METTLER-TOLEDO AG **CRN:** 0F12048.51**MDMT:** 32 deg F**MAWT:** 284 deg F**MAWP:** 232 psig**Drawing #:** Per Product List Scope Sheets**Drawing Revision:** 1.4**Conditions Of Registration:**

Renewal - Housing InFit 761 & InTrac 777 w/Sensors per att'd scope of registration sheets (4 pages).

This design was registered based on a technical review performed by the province of initial registration in accordance with the Association of Chief Inspectors policy on reciprocal recognition of design review.

Reviewer's Notes:

As required by CSA B51 4.2.1, this registration expires on February 20, 2028. This CRN is valid until the expiry date as long as the Manufacturer maintains a valid quality control program verified by an acceptable third-party agency until that date. Should the certification of the quality control program lapse before the expiry date, this registration shall become void.

Contact me if you have any questions. The invoice for registration will be forwarded under separate cover.

SHARON PETERS

boiler.designregistration@technicalsaftybc.ca
Design Administration**cc:**