



345 Carlingview Drive
Toronto, Ontario M9W 6N9
Tel.: 416.734.3300
Fax: 416.231.1626
Toll Free: 1.877.682.8772

www.tssa.org

May 26, 2020

CYNTHIA FORMANIUK
ALBERTA BOILERS SAFETY ASSOCIATION
9410 20 AVE NW
EDMONTON AB T6N 0A4
CA

Service Request Type: BPV-Fitting Registration
Service Request No.: 2846814
Your Reference No.: 2020-02161 - ABSA PORTAL #2020-S0160 CRN#0F10611.2
Registered to: KYTOLA INSTRUMENTS OY

Dear CYNTHIA FORMANIUK,

Technical Standards and Safety Authority (TSSA) is pleased to inform you that your submission has been reviewed and registered as follows:

CRN No.: 0F10611.25R1

Main Design No.: Scope of Registration - Variable Area Flowmeters
Expiry Date: 06-May-2030

Please be advised that a valid quality control system must be maintained for the fitting registration to remain valid until the expiry date.

The stamped copy of the approved registration and the invoice are mailed separately. Should you have any questions or require further assistance, please contact a Customer Service Advisor at 1.877.682.TSSA (8772) or e-mail customerservices@tssa.org. We will be happy to assist you. When contacting TSSA regarding this file, please refer to the Service Request number provided above.

Yours truly,

Liliana Constantinescu, P Eng
Tel. : 416-734-3425
lconstantinescu@tssa.org

STATUTORY DECLARATION
Registration of Fittings
Single or Multiple Fitting Designs within one Fitting Category



I, Sakari Häyrynen, President
(name of applicant) (position title) (must be in a position of authority)

of Kytola Instruments Oy
(name of manufacturer)

located at Olli Kytölän tie 1, 40950 Muurame, Finland
(plant address)

do solemnly declare that the fittings listed hereunder, which are subject to the Safety Codes Act (select only one)

- comply with the requirements of _____ which specifies the dimensions, (title of recognized North American Standard) materials of construction, pressure/temperature ratings and identification marking of the fittings, or
- are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with ASME B31.3 last edition as supported by the (title of code of construction or other applicable document) attached data which identifies the dimensions, materials of construction, pressure/temperature ratings and the basis for such ratings, and the identification marking of the fittings.

I further declare that the manufacture of these fittings is controlled by a quality control program which has been verified as described in the below Table as being suitable for the manufacturing of these fittings to the stated standard, regulation, code, guideline or other applicable document. The fittings covered by the declaration for which I seek registration are as provided in the Supplementary Sheet(s) attached.

Quality Program Verification and Manufacturing Sites

A copy of the Quality Certificate from each manufacturing site must be included

Item #	Product Description, Model or Series	Quality Program	Scope of Certification	Expiry Date	Verifying Organization	Location(s) Plant Name and address
1.	Constant Flow Regulators 2851 all models 2914 all models	ISO 9001	Design, research and development, manufacturing, sales and marketing of flow and control devices and analyzers.	30 September 2021	Det Norske Veritas	Kytola Instruments Oy Olli Kytölän tie 1, 40950 Muurame, Finland

In support of this application, the following information, calculations and/or test data are attached:

FEA & Drawings, data sheets, ISO 9001 Certificate,

Ismo Sakari Häyrynen
(Signature of the Declarer)

April 24, 2020
(Date)

DECLARED before me at Jyväskylä in the country of Finland
(city) (province, territory, or state)

this 24th day of April, 2020
(Month) (Year)

(print) _____
(a Commissioner of Oaths or Notary Public)

(sign) _____
(a Commissioner of Oaths or Notary Public)

(expiry date (mm/dd/yy))

This is to certify the own hand signature(s) of
Ismo Sakari Häyrynen
Jyväskylä, the 24th day of April 2020
Ex officio:

Fee EUR 9€

ANU MYLLYMÄKI
Henkikirjoittaja, julkinen notaari
District Registrar, Notary Public



Commissioner of Oaths / Notary Public in and for: Jyväskylä, Finland
(province, territory, or state)

For ABSA Office Use Only:

NOTES: _____

To the best of my knowledge and belief, the application meets the requirements of the Safety Codes Act and CSA Standard B51, Part 1, Clause 4.2, and is accepted for registration in Category F.

CRN: OF10611.25R1

Registered Date: MAY 26, 2020

Expiry Date: MAY 6, 2030

Signature: *[Signature]*
(Signature of the Administrator/SCO)

The information you provide is necessary only for the administration of the programs as required by the Alberta Safety Codes Act and Regulations in the Pressure Equipment Discipline

Technical Standards and Safety Authority
Boilers and Pressure Vessels Safety Program

REGISTERED

C.R.N.: OF10611.25R1

Signed: *[Signature]*

Date: May 26, 2020.

THIS IS A RENEWAL FOR SCOPE OF REGISTRATION SEE ATTACHED STAMPED DOCUMENT



Date : 16/04/2020

REFERENCE : OF10611.2 RENEWAL

- Scope of Registration -

Desin Code: ASME B31.3 last edition

KYTOLA Variable Area Flow Meter	MAWP	MAWT	MDMT
all models of VEH, VDH, and VLH	266 psig	248°F	-20°F
all models of LR & LH	239 psig	176°F	-20°F

Thanks and BR

The MAWT for model LH is 167 deg F

Martin Valois
Sales Engineer
(514) 691-9945
Martin.Valois@kytola.ca

THIS IS PART OF CRN
OF10611.25R1
Technical Standards and Safety Authority
Boilers and Pressure Vessels Safety
Program