

February 16, 2019

**Attention:** Cung Nguyen  
N-CUBED ENGINEERING  
8 MAIN STREET EAST SUITE 409  
HAMILTON, ON L8N 1E8

**Email:** CNguyen@ncubed.ca

The design submission, tracking number 2018-08376, originally received on December 18, 2018 was surveyed and accepted for registration as follows:

**CRN :** 0H16882.52 **Accepted on:** February 16, 2019  
**Reg Type:** NEW DESIGN **Expiry Date:** May 09, 2024  
**Drawing No. :** 00095479 INDEX A  
**Fitting type:** SENSOR

Design registered in the name of : METTLER TOLEDO GMBH, PROCESS ANALYTICS

Description	MAWP	Design Temperature	MDMT
Internal Pressure	1600kPa	130 °C	-29 °C
External Pressure	0kPa	130 °C	-29 °C

**The registration is conditional on your compliance with the following notes:**


*As indicated on AB-41 Statutory Declaration form and submitted documentation, the code of construction is ASME B31.3.*

- *This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration form.*
- *This registration is valid only until the indicated expiry date and only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date.*
- *Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.*

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

If you have any question don't hesitate to contact me by phone at (780) 433-0281 ext 3330 or fax (780) 437-7787 or e-mail grynchuk@absa.ca.

Sincerely,



GRYNCHUK, MILLA, P. Eng.  
DOP Cert. No. D00005217